

TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518
Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration

Certificate No TSP-06012022-009

Gage ID PEA-11515-2 Gage S/N 11515-2 Description Profile Gauge Operating Procedure: OP-129 Unit of Meas. Manufacturer Lone star Cal. Date 6/1/2022 Next Due 6/1/2025 Cal. Freq. 3.00 Years Location Lab	Environmental Conditions Temperature 68 +/- 2 deg F Humidity 20-55% Approved Yes Customer Info. Peak NDT
--	---

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 10638 JOB 11515

AS FOUND: PASSED AS LEFT: PASSED

Gage: NC-23-50, 4 TPI, 2 TPF

Profile Overlay Serial #: C14281
Profile Overlay NIST# 03-7-1857

Optical Comparator w/ Digital Readout (14"), Mitutoyo, Model PH-A14
Serial #: 012721312
Traceable to NIST #: Cert# CK324-29269-539
Last / Next Cal Dates: 03/04/2021 --> 03/04/2023

APPROVED*By Kayla Myers at 5:15 pm, Jun 03, 2022*Calibrated By Taichi Daimo

Signature

Date: 6/1/2022

TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518
Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration

Certificate No TSP-06012022-010

Gage ID PEA-11515-3 Gage S/N 11515-3 Description Profile Gauge Operating Procedure: OP-129 Unit of Meas. Manufacturer Lone star Cal. Date 6/1/2022 Next Due 6/1/2025 Cal. Freq. 3.00 Years Location Lab	Environmental Conditions Temperature 68 +/- 2 deg F Humidity 20-55% Approved Yes Customer Info. Peak NDT
--	---

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 10638 JOB 11515

AS FOUND: PASSED AS LEFT: PASSED

Gage: PG-542, 4 TPI, 2 TPF, V.050

Profile Overlay Serial #: C14282

Profile Overlay NIST# 03-7-1857

Optical Comparator w/ Digital Readout (14"), Mitutoyo, Model PH-A14

Serial #: 012721312

Traceable to NIST #: Cert# CK324-29269-539

Last / Next Cal Dates: 03/04/2021 --> 03/04/2023

APPROVED**By Kayla Myers at 5:16 pm, Jun 03, 2022**Calibrated By Taichi Daimo

Signature

Date: 6/1/2022

TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518
Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration

Certificate No TSP-06012022-011

Gage ID PEA-11515-4 Gage S/N 11515-4 Description Profile Gauge Operating Procedure: OP-129 Unit of Meas. Manufacturer Lone star Cal. Date 6/1/2022 Next Due 6/1/2025 Cal. Freq. 3.00 Years Location Lab	Environmental Conditions Temperature 68 +/- 2 deg F Humidity 20-55% Approved Yes Customer Info. Peak NDT
--	---

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 10638 JOB 11515

AS FOUND: PASSED AS LEFT: PASSED

Gage: NC-56-77, 4 TPI, 3 TPF

Profile Overlay Serial #: C14283
Profile Overlay NIST# 03-7-1857

Optical Comparator w/ Digital Readout (14"), Mitutoyo, Model PH-A14
Serial #: 012721312
Traceable to NIST #: Cert# CK324-29269-539
Last / Next Cal Dates: 03/04/2021 --> 03/04/2023

APPROVED**By Kayla Myers at 5:16 pm, Jun 03, 2022**Calibrated By Taichi Daimo

Signature



Date:

6/1/2022

TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518
Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration

Certificate No TSP-06012022-012

Gage ID PEA-11515-5 Gage S/N 11515-5 Description Profile Gauge Operating Procedure: OP-129 Unit of Meas. Manufacturer Lone star Cal. Date 6/1/2022 Next Due 6/1/2025 Cal. Freq. 3.00 Years Location Lab	Environmental Conditions Temperature 68 +/- 2 deg F Humidity 20-55% Approved Yes Customer Info. Peak NDT
--	---

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 10638 JOB 11515

AS FOUND: PASSED AS LEFT: PASSED

Gage: PG-543, 4 TPI, 3 TPF, V.050

Profile Overlay Serial #: C14284
Profile Overlay NIST# 03-7-1857

Optical Comparator w/ Digital Readout (14"), Mitutoyo, Model PH-A14
Serial #: 012721312
Traceable to NIST #: Cert# CK324-29269-539
Last / Next Cal Dates: 03/04/2021 --> 03/04/2023

APPROVED**By Kayla Myers at 5:17 pm, Jun 03, 2022**Calibrated By Taichi Daimo

Signature

Date: 6/1/2022

TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518
Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration

Certificate No TSP-06012022-013

Gage ID PEA-11515-6 Gage S/N 11515-6 Description Profile Gauge Operating Procedure: OP-129 Unit of Meas. Manufacturer Lone star Cal. Date 6/1/2022 Next Due 6/1/2025 Cal. Freq. 3.00 Years Location Lab	Environmental Conditions Temperature 68 +/- 2 deg F Humidity 20-55% Approved Yes Customer Info. Peak NDT
--	---

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 10638 JOB 11515

AS FOUND: PASSED AS LEFT: PASSED

Gage: PG-453, 5 TPI, 3 TPF, V.040

Profile Overlay Serial #: C14285
Profile Overlay NIST# 03-7-1857

Optical Comparator w/ Digital Readout (14"), Mitutoyo, Model PH-A14
Serial #: 012721312
Traceable to NIST #: Cert# CK324-29269-539
Last / Next Cal Dates: 03/04/2021 --> 03/04/2023

APPROVED*By Kayla Myers at 5:18 pm, Jun 03, 2022*Calibrated By Taichi Daimo

Signature

Date: 6/1/2022