

TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518
 Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration**Certificate No TSP-04292025-013**

| | |
|--|---|
| Gage ID PEA-11515-1 Gage S/N 11515-1 Description 20-0-20 Gauss Meter Operating Procedure: OP-117 Unit of Meas. Gauss Manufacturer R. B. Annis Co., Inc. Cal. Date 4/29/2025 Next Due 10/29/2025 Cal. Freq. 6.00 Months Location Lab | Environmental Conditions Temperature 68 +/- 2 deg F Humidity 20-55% Approved Yes Customer Info. Peak NDT |
|--|---|

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

| | | | | | | | |
|----------------------------|-------------|--------------------|---------|-----------------|--------|--------------------|----|
| Standard ID | +10 | Uncertainty | | Units | | Type | V |
| Limited Use? | No | Minimum | 9 | Nominal | 10 | Maximum | 11 |
| Ref Type | Gauss Meter | As Found | 10 | Accuracy | 0 | Fail Before | No |
| | | As Left | 10 | Accuracy | 0 | Fail After | No |
| Gage ID of Standard | Ref-09-1229 | | | | | | |
| Std Due Date | 10/10/2025 | Gage S/N | 09-1229 | NIST No. | 114537 | | |
| Standard ID | +20 | Uncertainty | | Units | | Type | V |
| Limited Use? | No | Minimum | 19 | Nominal | 20 | Maximum | 21 |
| Ref Type | Gauss Meter | As Found | 20 | Accuracy | 0 | Fail Before | No |
| | | As Left | 20 | Accuracy | 0 | Fail After | No |
| Gage ID of Standard | Ref-09-1229 | | | | | | |
| Std Due Date | 10/10/2025 | Gage S/N | 09-1229 | NIST No. | 114537 | | |
| Standard ID | -10 | Uncertainty | | Units | | Type | V |
| Limited Use? | No | Minimum | 9 | Nominal | 10 | Maximum | 11 |

| | | | | | | | |
|----------|-------------|----------|----|----------|---|-------------|----|
| Ref Type | Gauss Meter | As Found | 10 | Accuracy | 0 | Fail Before | No |
| | | As Left | 10 | Accuracy | 0 | Fail After | No |

Gage ID of Standard Ref-09-1229

Std Due Date 10/10/2025

Gage S/N 09-1229

NIST No. 114537

| | | | | | | | |
|--------------|-------------|-------------|----|----------|----|-------------|----|
| Standard ID | -20 | Uncertainty | | Units | | Type | V |
| Limited Use? | No | Minimum | 19 | Nominal | 20 | Maximum | 21 |
| Ref Type | Gauss Meter | As Found | 20 | Accuracy | 0 | Fail Before | No |
| | | As Left | 20 | Accuracy | 0 | Fail After | No |

Gage ID of Standard Ref-09-1229

Std Due Date 10/10/2025

Gage S/N 09-1229

NIST No. 114537

Calibrated By Taichi Daimo

Signature



Date: 4/29/2025

APPROVED*By Kayla Myers at 2:50 pm, Apr 30, 2025*