# TSP Calibration, Inc.

1110 S. FM 1788 Suite 38 Midland, TX 79706 Phone: (432) 296-4992 Email: Chad@tspndt.com

## **Certificate of Calibration**

### Certificate No TSPW-05162025-001

Gage ID PEA-9-622-1

Gage S/N 9-622-1

Description 20-0-20 gauss meter

Operating

OP-117

Procedure:

Unit of Meas. gauss

Manufacturer R.B. Annis

Cal. Date 5/16/2025

Next Due 11/16/2025

Cal. Freq. 6.00

Months

Location Lab

**Environmental Conditions** 

Temperature

68 Deg F +/- 2

Humidity

35-55%

Approved Yes

Customer Info. Peak NDT Solutions

#### **Certification Statement**

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

#### **Findings**

SO 14313 Job 14161

Standard ID	+10		Uncertainty			Units			Туре	V	
Limited Use?	No		Minimum		9	Nominal		10	Maximum	11	
Ref Type	20-0-2	20 Gauss Meter	As Found		10	Accuracy		0	Fail Before	No	
			As Left		10	Accuracy		0	Fail After	No	
Gage ID of Sta	ndard	REF-TSPSTD-02									
Std Due Date		8/8/2025	Gage S/N	TSPSTD-02		NIST No.	154784				
Standard ID	+20	, *	Uncertainty			Units	4	-	Туре	V	1
Limited Use?	No		Minimum		19	Nominal		20	Maximum	21	
Ref Type	20-0-2	20 Gauss Meter	As Found		20	Accuracy		0	Fail Before	No	
			As Left		20	Accuracy		0	Fail After	No	
Gage ID of Star	ndard	REF-TSPSTD-02									
Std Due Date		8/8/2025	Gage S/N	TSPSTD-02	1	NIST No.	154784				
Standard ID	-10	10 5 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1	Uncertainty			Units	who are construction to the construction of th		Туре	V	
Limited Use?	No		Minimum		9	Nominal		10	Maximum	11	

Fail Before No 0 As Found 10 Accuracy 20-0-20 Gauss Meter Ref Type Fail After No 0 10 Accuracy As Left Gage ID of Standard REF-TSPSTD-02 NIST No. 154784 Gage S/N TSPSTD-02 8/8/2025 Std Due Date Type V Units Uncertainty Standard ID -20 20 Maximum 21 19 Nominal Limited Use? No Minimum 0 Fail Before No As Found 20 Accuracy 20-0-20 Gauss Meter Ref Type Fail After No 0 As Left 20 Accuracy Gage ID of Standard REF-TSPSTD-02

Gage S/N TSPSTD-02

Calibrated By

Std Due Date

Gage ID: PEA-9-622-1

**Chad Adams** 

8/8/2025

**Signature** 

NIST No.

154784

Date:

Page

5/16/2025

2 of 2

# **APPROVED**

By Kayla Myers at 4:04 pm, May 19, 2025