TSP Calibration, Inc.

1110 S. FM 1788 Suite 38-40 Midland, TX 79706 Phone: (432) 307-7054 Email: cory@tspndt.com

Certificate of Calibration

Certificate No TSPW-02212022-008

Gage ID PEA-25437

Gage S/N 25437

Description 10 lb Test Weight

Operating OP-118

Procedure:

Unit of Meas. pounds Manufacturer Parker

> Cal. Date 2/21/2022 Next Due 2/21/2025

Cal. Freq. 3.00

Location Lab

Years

Environmental Conditions

Temperature

68 Deg F +/- 2

Humidity

35-55%

Approved Yes

Customer Info. Peak NDT Solutions

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

AS FOUND: PASSED

AS LEFT: PASSED

Using calibrated scale serial no. 8335029565

Test weight as found 10.110 lbs.

Test weight as left 10.110 lbs.

Calibration due date for scale is on 06/29/2022.

Scale is traceable to NIST by the following numbers: Doc# 109087

Calibrated By Chad Adams

Signature

Date:

2/21/2022

APPROVED

By Kayla Myers at 2:10 pm, Mar 09, 2022