

### TSP Calibration, Inc.

1110 S. FM 1788 Suite 38-40 Midland, TX 79706  
Phone: (432) 307-7054 Email: Chad@tspndt.com

## Certificate of Calibration

Certificate No TSPW-10112021-016

<p><b>Gage ID</b> PEA-1011-4  <b>Gage S/N</b> 1011-4  <b>Description</b> Profile Gauge  <b>Operating Procedure:</b> Visual  <b>Unit of Meas.</b>  <b>Manufacturer</b> PMC Lone Star  <b>Cal. Date</b> 10/11/2021  <b>Next Due</b> 10/11/2024  <b>Cal. Freq.</b> 3.00            Years  <b>Location</b> Lab</p>	<p><b>Environmental Conditions</b>  Temperature 68 Deg F +/- 2  Humidity 35-55%</p> <p><b>Approved</b> Yes  <b>Customer Info.</b> Peak NDT Solutions</p>
--	--

### Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

### Findings

Flatness- PASS  
Damage- PASS  
Wear- PASS

As Found: Passed  
As Left: Passed

Calibrated By Chad Adams      Signature       Date: 10/11/2021

**APPROVED**  
By Kayla Myers at 1:27 pm, Jul 21, 2022

### TSP Calibration, Inc.

1110 S. FM 1788 Suite 38-40 Midland, TX 79706  
Phone: (432) 307-7054 Email: Chad@tspndt.com

## Certificate of Calibration

Certificate No TSPW-10112021-015

<p><b>Gage ID</b> PEA-1011-3  <b>Gage S/N</b> 1011-3  <b>Description</b> Profile Gauge  <b>Operating Procedure:</b> Visual  <b>Unit of Meas.</b>  <b>Manufacturer</b> PMC Lone Star  <b>Cal. Date</b> 10/11/2021  <b>Next Due</b> 10/11/2024  <b>Cal. Freq.</b> 3.00      Years  <b>Location</b> Lab</p>	<p><b>Environmental Conditions</b>  <b>Temperature</b> 68 Deg F +/- 2  <b>Humidity</b> 35-55%</p> <p style="text-align: center;"><b>Approved</b> Yes  <b>Customer Info.</b> Peak NDT Solutions</p>
--	--

### Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

### Findings

Flatness- PASS  
Damage- PASS  
Wear- PASS

As Found: Passed  
As Left: Passed

Calibrated By Chad Adams      Signature       Date: 10/11/2021

**APPROVED**  
By Kayla Myers at 1:27 pm, Jul 21, 2022

### TSP Calibration, Inc.

1110 S. FM 1788 Suite 38-40 Midland, TX 79706  
Phone: (432) 307-7054 Email: Chad@tspndt.com

## Certificate of Calibration

Certificate No TSPW-10112021-014

<p><b>Gage ID</b> PEA-1011-2  <b>Gage S/N</b> 1011-2  <b>Description</b> Profile Gauge  <b>Operating Procedure:</b> Visual  <b>Unit of Meas.</b>  <b>Manufacturer</b> PMC Lone Star  <b>Cal. Date</b> 10/11/2021  <b>Next Due</b> 10/11/2024  <b>Cal. Freq.</b> 3.00      Years  <b>Location</b> Lab</p>	<p><b>Environmental Conditions</b></p> <p>Temperature 68 Deg F +/- 2  Humidity 35-55%</p> <p style="text-align: center;"><b>Approved</b> Yes  <b>Customer Info.</b> Peak NDT Solutions</p>
--	--

### Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

### Findings

Flatness- PASS  
Damage- PASS  
Wear- PASS

As Found: Passed  
As Left: Passed

Calibrated By Chad Adams      Signature       Date: 10/11/2021

**APPROVED**  
By Kayla Myers at 1:28 pm, Jul 21, 2022

### TSP Calibration, Inc.

1110 S. FM 1788 Suite 38-40 Midland, TX 79706  
Phone: (432) 307-7054 Email: Chad@tspndt.com

## Certificate of Calibration

Certificate No TSPW-10112021-013

<p><b>Gage ID</b> PEA-1011-1  <b>Gage S/N</b> 1011-1  <b>Description</b> Profile Gauge  <b>Operating Procedure:</b> Visual  <b>Unit of Meas.</b>  <b>Manufacturer</b> PMC Lone Star  <b>Cal. Date</b> 10/11/2021  <b>Next Due</b> 10/11/2024  <b>Cal. Freq.</b> 3.00      Years  <b>Location</b> Lab</p>	<p><b>Environmental Conditions</b></p> <p>Temperature    68 Deg F +/- 2  Humidity        35-55%</p> <p style="text-align: center;"><b>Approved</b> Yes  <b>Customer Info.</b> Peak NDT Solutions</p>
--	--

### Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

### Findings

Flatness- PASS  
Damage- PASS  
Wear- PASS

As Found: Passed  
As Left: Passed

Calibrated By Chad Adams      Signature       Date: 10/11/2021

**APPROVED**  
By Kayla Myers at 1:28 pm, Jul 21, 2022